

महाराष्ट्र शासन  
इंदिरा गांधी शासकीय वैद्यकीय महाविद्यालय व रुग्णालय, नागपुर  
महात्मा ज्योतीराव फुले जन आरोग्य योजना

सेंट्रल एवेन्यु रोड, नागपुर - 18

Contact No. : 8275094649

email : [iggmchospital@gmail.com](mailto:iggmchospital@gmail.com),

क्रमांक इंगाशावैमवरुना / मज्योफुजआयो / 1426 / 2024

दि. 2/9 / 2024

## जाहिरात

इंदिरा गांधी शासकीय वैद्यकीय महाविद्यालय व रुग्णालय  
नागपुर  
महात्मा ज्योतीराव फुले जन आरोग्य योजना

इंदिरा गांधी शासकीय वैद्यकीय महाविद्यालय व रुग्णालय नागपुर येथील महात्मा ज्योतीराव फुले जन आरोग्य योजना अंतर्गत येणाऱ्या गरीब रुग्णांसाठी, कॉक्लेअर इंप्लान्ट कान नाक घसा विभागाला अत्यंत आवश्यकता असते. पुरवठ्यासाठी पुरवठादारांकडून दर करारा करीता दरपत्रके मागविण्यात येत आहे. याबाबत संपुर्ण माहिती इंदिरा गांधी शासकीय वैद्यकीय महाविद्यालय व रुग्णालया च्या संकेत स्थळावर [www.iggmc.org](http://www.iggmc.org) दि. 2/9/2024 ला उपलब्ध होईल.

*Acharya*

अधिष्ठाता

इंदिरा गांधी शासकीय वैद्यकीय महाविद्यालय  
व रुग्णालय नागपुर




## Terms & Conditions

1. Please quote your prices for each item on or before date 11 / 9 / 24 upto 5.00 PM. in a sealed envelope. Unsealed or improperly sealed quotation will not be accepted. Quotation received after 5.00pm will not be accepted.
2. Mention enquiry number of this letter on sealed quotation envelope & submit that in the name of "Dean IGGMC, Nagpur (MJPJAY) between 11.00 AM to 5.00 PM at the In word / Out word office of, IGGMC Nagpur. Sealed envelope will be opened at 12.00 PM on \_\_\_\_\_ at Dean's office IGGMC.
3. Your terms & conditions if any for supply should be mentioned in your covering letter enclosed with the quotation. Two bid should be submitted 1<sup>st</sup> is Technical & 2<sup>nd</sup> is price bid .
4. The serial number of the items should not be changed while quoting rates. You may drop the item if not interested & list should be prepared in printed copy only otherwise it may be rejected.
5. Submit the attested copy of
  - i) TIN / TAN / GST Number
  - ii) Pan Card
  - iii) Gumastha for 3 years
  - iv) Last year ITR
  - v) Bank Account Details For NEFT/ RTGS
6. Supplied items should be of good quality, ISI Marked & as per FDA norms. Mention proper specification of goods while quoting the prices.
7. If product will not be used and expired it will be return to supplier before 1 month of expiry date and supplier should have to return that product of new batch & expiry.
8. Please note that prices should inclusive of all taxes. Supplier should supply all goods as per these rates for 1 year or till the date of next enquiry whichever is later.
9. Delivery of goods will be accepted at pharmacy section in MJPJAY office. You have to execute supply order in stipulated period only. If not, supply order will be treated as cancelled.



10. Suppliers should mention exact name, rate & serial no of item in DM as given in quotation.
11. As the DM is certified, confirmatory order of respective DM from Hon' Dean IGGMC will be issued to supplier. Supplier should submit the tax invoice for respective confirmatory order at MJPJAY office within 3 days otherwise payment will be cancelled. After submission of tax invoice within 3 days payment will be processed within 45 days from date of bill submission.
12. In case of deficiency in service/quality the supplier will be debarred after three complaints for one year & supplier with second lowest rate will be selected for supply.
13. This office reserve the right to cancel the order at any time without showing any reason.



Dean

Indira Gandhi Government Medical College  
& Hospital Nagpur (MJPJAY)



PRODUCT SPECIFICATION COMPLIANCE

Sr.No.	Product	Desired specification of TIA
1	Implant	Titanium casing of non-Ceramic material
2	Intra-Cochlear Electrode Array	1.Multi-Channel /Multi Electrode with minimum of 12 electrode 2.Minimum length of active electrode - 25mm 3.Maximum diameter at the apex - 0.5
3	Model	Straight Slim Electrode array
4	Speech Processor	1.Advanced SP with option of Body worn and BTE 2.Multi-speech coding strategy capability utilizing digitally programmed frequency band (20 or more) 3.Should have direct streaming with phone without any intermediate device 4.Should have control of sensitivity, volume, alarm & auto features 5.Wide range of input processing option for clarity of sound in different environments like soft sound, Noise, Music and focus listening 6.Protection against humidity & sweat 7. External speech processor of Cochlear Implant should have single/dual microphone facility 8.Mobile app application to monitor speech, battery life & " find my device" to find misplaced processor.
5	Power source	Powered by rechargeable cells / rechargeable battery packs
6	Implant Integrity Testing ability	1.Ability to test implant and electrode integrity and function both intraoperative and postoperatively 2.Feature for testing intra op implant functioning 3.Inner ear assessment at the time of surgery and during switch on mapping & therapy after CI surgery 4.Company should provide one working equipment for automatic intra operative telemetry/compound action potential recoding during entire time of the tender free of cost.
7	Speech Coding strategy	Proven and internationally accepted speech coding strategy
8	MRI compatibility	1.MRI compatibility of 1.5 Tesla or more without removal of magnet. 2.Implant should preferably have option of removable magnet for 3.0 T or higher MRI compatibility.
9	FDA Approval	Approved Complete implant system should be US FDA for implantation in paediatric patients
10	Feature of Implant	1.Implant should be compatible with current and future technology in speech processor. 2.If company is no longer manufacturing the current processor, then it should be upgraded to higher model free of cost if within warranty period of internal device
11	External components	1.Robust and long lasting. 2.Should be sweat/splash resistant. 3.Easily available spare parts like magnet coil, wire and microphone sound processor etc
12	FM compatibility	Compatible with standard FM bases group Therapy Systems / FM



13	Telephone / Telecoil compatibility	Telephone compatibility with inbuilt snap/ on Telephone Adapter.
14	Product Support	1.For the life of the product including implanted and external components, software and hardware 2.Free upgrade of software/ hardware support for product in case existing implant system is discontinued by manufacturer
15	Infrastructure support	All necessary hardware and software required for preoperative intra-operative and post operative testing and programming of the Implant should be provided free of cost to existing implant centres and any centres established in the future.
16	Service network	1.Provision of spares/loaners for equipment defects/ malfunctions Within 48 hours of request for such support 2.Should have wide network of service and repair centres all over India to provide after sales service like repair/replacement within India of the warrantable components which includes wire, microphone, sound processor, any other accessories like magnet coil etc. to the patients residing in various parts of India, directly by the company
17	Comprehensive replacement / repair warranty	1.Implant - 10 yrs from date of surgery 2.Processor- 3 yrs from date of switch -on 3.Accessories - 1 yr (for battery ,cable, coil etc.)
18	Provision for return	Company should have provision of return of implant in case non utilization of implant for any reason.
19	Kit Inclusions	1.Coil and cable - 2 Qty. 2. Rechargeable batteries- 2 Nos. 3.Charger for rechargeable battery. 4.Monitor earphones. 5.Magnets of varying strengths. 6.Disposbale batteries.- 1 Box
20	Reliability	Implant should have good reliability report as per ANSI/AAMI European standards.
21	Provision of surgical instrument set	Company should provide implant specific surgical instrument set for smooth conduct of surgical procedure.(2 Qty. prefer)
22	Parent training & counselling	Regular training & counselling of parents and care givers regarding the maintenance of speech processor.

*K. K. K.*  
for HOD ENT.

*Acharya*  
अधिष्ठाता  
इंदिरा गांधी शासकीय वैद्यकीय महाविद्यालय  
व रुग्णालय, नागपूर.